



AMNIO SWAB

AMNIO SWAB is a convenient swab screening system intended as an aid in detecting rupture of the amniotic membrane in pregnant women.

RELIABLE

- Quality system certified ISO 9001 and ISO 13485.
- Sterilized swabs by gamma radiation.

PRACTICAL

- The swabs are packaged in individual sleeves.
- Positive controls available upon request (pH : 5.0–5.5–6.0–6.5–7.0–7.5).
- Storage between 15–30°C.
- Long shelf life : 36 months.

FAST

- Immediate results.
- Visual reading.

PRESENTATION

- Kit of 100 tests contains :
 - 100 swabs in a single sterilizes sleeve,
 - 1 colour card.

Rupture of the amniotic membrane can result in small volumes of amniotic fluid leaking into the upper vagina. The presence of amniotic fluid tends to elevate the pH of the upper vagina. Detection of this pH increase using a pH indicator dye has been shown to assist in determining the presence of amniotic fluid.



Ekoweb Ltd.

Kaarina
Finland

Tel. +358 50 3633969

Fax +358 22433786

info@ekoweb.fi

www.ekoweb.fi

VAT ID

FI10309538

- 1 - Remove a swab from its protective sleeve. DO NOT touch the tip of the swab or allow it to come into contact with any liquid or other substance which might affect pH.
- 2 - Using a vaginal speculum, part the labia exposing the cervix and carefully insert the swab into the vagina. Do not allow the swab to come into contact with vaginal tissue during entry.
- 3 - Allow first and only contact of the **AMNIOTEST**[®] swab tip to occur with upper vaginal tissue (posterior vaginal fornix and external cervical os).
- 4 - Allow the tip to remain in contact with upper vaginal tissue for about 15 seconds.
- 5 - Carefully remove the swab and immediately examine the color of the tip.

