



Report

Evaluation of system accuracy of the blood glucose monitoring system “ACON
On Call Sure” according to DIN EN ISO 15197: 2015

Sponsor:	ACON Laboratories Inc. 10125 Mesa Rim Road San Diego, Ca92121 USA
The study was conducted in:	Institute of Diabetes „Gerhardt Katsch“ Karlsburg e.V. Greifswalder Straße 11 e 17495 Karlsburg
Study director:	Dr. Eckhard Salzsieder Institute of Diabetes „Gerhardt Katsch“ Greifswalder Str. 11 E 17495 Karlsburg
Author:	Dr. Sabine Berg Dr. E.-J. Freyse
Date of realization:	September / October 2018

1.	Introduction	3
2.	Aim	3
3.	Study design	3
3.1	Test persons.....	3
3.2.	Study period	4
3.3.	Screening.....	4
3.4.	Risks of experimental procedure / termination criteria.....	4
3.5.	Material and methods.....	4
3.5.1	Test device.....	4
3.5.2	Reference device.....	5
3.5.3	Determination of hematocrit	6
4.	Execution of the study in accordance with the guidelines of DIN EN ISO 15197:2013.....	6
4.1.	Samples	6
4.2.	Blood glucose measurement systems in the laboratory.....	7
4.3.	Environmental conditions.....	7
4.4.	Additional exclusion criteria	7
4.5.	Determination of glucose concentration using test strips.....	7
5.	Analysis / Results.....	8
5.1.	Time of realization.....	8
5.2.	Data	8
5.3.	Analysis of system accuracy in accordance with DIN EN ISO 15197:2013..	8
6.	Summary and conclusions	13
	References	13
	Signatures.....	13

1. Introduction

The aim of modern diabetes therapy is the maintenance of near normal blood glucose concentrations to prevent both hypo- and hyperglycaemic excursions that have been closely associated with the development of acute and late complications of the disease.

Blood glucose monitoring/measurement systems, which can be used for self-determination of the blood glucose by the patient, are helpful for a better metabolic control by the patients themselves and by the physician and allow a more flexible adjustment of the individual medication/therapy.

To achieve this goal, any insulin dose adjustment should precede the blood glucose self control using blood glucose monitoring systems with a high quality and measurement accuracy.

The directive DIN EN ISO 15197 [1] demands strict international quality standards for the system accuracy of blood glucose monitoring systems (BGMS). The DIN EN ISO 15197:2015 [2] represents a revision of the DIN EN ISO 15197:2003 [1]. The minimal acceptable accuracy for measurements using BGMS is defined as follows:

For blood glucose concentrations < 5.55 mmol/l (< 100 mg/dl) the glucose values are allowed to differ in comparison to the reference method up to ± 0.83 mmol/l (± 15 mg/dl).

In the case of blood glucose concentrations of ≥ 5.55 mmol/l (≥ 100 mg/dl) only deviations of less than 15% as compared to the reference method are acceptable.

In addition, the above criteria must be met by at least 95 % of the measured values.

According to § 23b MPG, the present study has tested the system accuracy in accordance to DIN EN ISO 15197:2015 for the system On Call Sure.

If systems without CE-marks are tested, the manufacturer/customer of the study must provide a Manufacturer's Declaration according to § 12 (3), § 24 (2) of the Medical Devices Act and Annex VIII of the Directive 98/79/EC of *in vitro* diagnostic medical devices of the European Parliament and of the Council of Oct. 27, 1998. This Manufacturer's Declaration will be submitted to the Health Authority in the context of an amendment.

The glucose monitoring system had a CE certification.

Volunteers were insured by an insurance agency.

2. Aim

The aim of this investigation was to proof the system accuracy of the On Call Sure in accordance with the directive DIN EN ISO 15197: 2015. Reference measurements were performed with the YSI 2300 STAT PLUS system.

3. Study design

Capillary blood, taken from the fingertip of patients (skin disinfection before puncture with a lancet), was used for the measurement of blood glucose concentration by the test- (On Call Sure) and reference method (YSI 2300).

3.1 Test persons

The volunteers had to meet the following inclusion criteria:

- Male or female patients with hypo-, normo- or hyperglycaemia
- The written informed consent had to be signed.
- The volunteers must be older than 18 years.
- The volunteers have legal capacity and are able to understand meaning, nature and possible consequences of the test.
- The subjects will receive 10.00 € for expenses.

The exclusion criteria:

- Pregnancy or lactation.
- Acute or chronic diseases with the risk of aggravation by the measure.
- A current constitution which does not allow participating in the study.
- By the patient exclusion criteria given in the user manual of the monitoring system were reached.

3.2. Study period

The blood withdrawal in a patient took 5 to 10 minutes.

3.3. Screening

To attract potential subjects for the study, peoples were first informed about the objective, procedure, risk and duration of the study. After declaration of willingness to participate in the study, the written consent from the volunteer was requested. On the experimental day, good physical fitness was a prerequisite for blood sampling.

3.4. Risks of experimental procedure / termination criteria

The blood sampling was executed by qualified persons under strict hygienic conditions, using only disposable material to minimize the risks for the subjects. In case of discomfort of a subject, blood sampling was interrupted.

3.5. Material and methods

3.5.1. Test device Blood Glucose Monitoring System **On Call Sure**

Manufacturer: ACON Laboratories Inc.
10125 Mesa Rim Road
San Diego, Ca92121
USA

Technical data:

Blood sample: capillary, venous, neonatal blood
Sample volume: 0.6 µl
Glucose measurement range: 10 – 600 mg/dl
Measuring time: about 5 sec
Working temperature: 5 – 45 °C
Relative humidity: 10 – 90 %
Hematocrit: 10 – 70 %
Technology: electrochemical dynamic
Enzyme: GDH - FAD
Calibration: plasma equivalent
Code: autocoding

For the tests, two blood glucose monitors On Call Sure were available and used during the tests.

Serial number and study code of the provided two monitors:

Serial number	Study code
SN404E43DA562	GC1
SN404E43DA557	GC2

Test strips

In total, 600 test strips from each of 3 lots were available. Reagent systems from 11 packs were used per lot. The following lots were included into the tests:

Test strips		
Numbering	Lot No.	Expiration date
Lot 1	1790007	06/2020
Lot 2	1790008	06/2020
Lot 3	1790009	06/2020

Control measurements

The control measurements were carried out using three control solutions, whose characteristics are listed in the following table:

Control solution	Lot No.	Expiration date	Target area (mg/dl) of test strips		
			Lot 1	Lot 2	Lot 3
Level High	SH80004	07/2020	277 – 416	273 – 411	278 - 417
Level Medium	SN80004	07/2020	88 – 133	87 – 131	90 - 136
Level Low	SL80004	07/2020	23 - 53	21 - 52	24 - 54

The measurements were performed according to the instructions by the manufacturer. On each study day before the test measurements, control measurements in each range and for each test strip lot and each monitor were performed (Excel table “controls”).

The system was found to be appropriate when the control measurements fell into the range that has been specified by the manufacturer.

The results of the control measurements were within the desired range, so that no device had to be replaced.

The measurements of the control solutions had in every case a deviation to the mean of less than 15 %.

3.5.2. Reference device

YSI 2300 STAT PLUS Manufacturer: YSI Incorporated, Yellow Springs, Ohio, USA

Method: Glucose Oxidase (GOD)

Probe: Lithium-Heparin-Plasma sampled from capillary blood

Duplicate = measurement with 2 electrodes

Sample volume: 25 µl

Glucose measurement range: up to 900 mg/dl (to 50 mmol/l)

Measuring time: 65 sec

Working temperature: 15 - 35 °C

Relative humidity: 10 - 90 %

Calibration: at 180 mg/dl

Accuracy and Precision

The proof of accuracy and precision was performed by use of YSI Standards with glucose concentrations of 90 mg/dl, 180 mg/dl, 450 mg/dl and 900 mg/dl, respectively. The standards were measured on the test day, before, during and after the test series.

Maintenance, adjustment and control procedures

For all the equipment used during the study, the control procedure has been implemented according to the manufacturer's instructions.

3.5.3. Determination of hematocrit

The procedure was performed in accordance with DIN 58933-1. The blood was sampled in heparinized micro-hematocrit capillaries (Laboratory Glassware, Marienfeld, Germany), which were closed and centrifuged thereafter (Hettich centrifuge HAEMATOKRIT 200; Germany). Reading the hematocrit value was performed by using a Nomogram.

4. Execution of the study in accordance with the requirements of DIN EN ISO 15197:2015

4.1. Samples

A total of 105 capillary blood samples were taken. After evaluation of glucose concentration ranges using the reference method YSI, exactly 100 samples were included into the stat. To reach the lower blood glucose ranges < 80 mg/dl, the glucose concentration was decreased by storage of the samples at 37 °C (stored).

To reach the higher blood glucose ranges, the Lithium-Heparin-blood (300 µl) was spiked with a 40 % glucose solution (B. Braun, Melsungen, Deutschland) (spiked).

For the measurements of original blood samples (original), 100 µl blood were sampled from the fingertip into Lithium-Heparin tubes (Mikrovetten[®], Sarstedt AG & Co., Nümbrecht, Germany) before and after each test measurement. These samples were used for reference measurements (YSI) before and after test strip measurements. In accordance with the requirements as described in DIN EN ISO 15197:2015, samples of the following concentration ranges were included into the procedure:

Range	Percentage of samples (%)	Glucose concentration(mg/dl)	Sample handling
1	5	≤ 50	all stored (5)
2	15	> 50 - 80	8 unchanged and 7 stored
3	20	> 80 - 120	all unchanged
4	30	> 120 - 200	all unchanged
5	15	> 200 - 300	all unchanged
6	10	> 300 - 400	5 unchanged and 5 spiked
7	5	> 400	all spiked (5)

The assignment to the concentration ranges based on the results of the reference measurements that have been done using the YSI 2300 STAT PLUS.

4.2. Blood glucose monitoring systems

Before start of the study, medical technical personnel was instructed in correct handling of the measuring systems. The devices had been properly maintained. Strips of each test lot were measured on two different devices.

4.3. Environmental conditions in the laboratory

During the study, room temperature ranged between 19.5 and 24.1 °C and the humidity between 41 and 57 %. (Excel table “conditions during the test”). So the prescribed external conditions were met for the blood glucose measurement systems (see characteristics of the test and reference device).

4.4. Additional exclusion criteria

The glucose concentration ranges were covered in accordance with the ranges in the table (see 4.1.) and unsuited blood samples (n=5) were excluded. Criterion for the exclusion was a difference of more than 4 % between reference value 1 and 2. A failure during the measurement was documented, or the range of measurement was already completed. The hematocrit of the samples should range between 10 - 70 % (a set in the manual of the manufacturer), which was observed/measured in all samples. Accordingly, there was no reason to exclude blood samples due to hematocrit abnormalities. Only those data were included if the coefficient of variation of the duplicates - determined using the reference method – was below 4 %.

4.5. Determination of glucose concentration using test strips

For use of original unchanged blood samples, capillary blood was taken directly from the fingertip to get contact to the test strip.

For reference measurements, blood was sampled into Lithium Heparin tubes (Mikrovetten[®]). The reference measurements using the YSI apparatus were performed in plasma samples.

Sampling was performed in the following order:

1. 100 µl blood were taken from the fingertip into Lithium Heparin tubes for reference measurement (R1).
2. BG measurements:
 - On Call Sure – GC 1 Lot 1
 - On Call Sure – GC 2 Lot 1
 - On Call Sure – GC 1 Lot 2
 - On Call Sure – GC 2 Lot 2
 - On Call Sure – GC 1 Lot 3
 - On Call Sure – GC 2 Lot 3
3. 100 µl blood were taken from the fingertip into Lithium Heparin tubes for reference measurement (R2)
4. Blood sampling for determination of the hematocrit.

To get samples in the different by ISO defined ranges, the glucose concentration of any taken sample was adjusted in accordance with the instructions of DIN EN ISO 15197:2015 (blood spiked or blood stored). The spiked blood samples were safety mixed by rotation for 30 minutes at room temperature. Thereafter, an aliquot was taken for YSI reference measurements before and after BGM testing.

To get glucose concentrations < 80 mg/dl, blood samples were incubated at 37 °C in a shaking water bath for 1 – 3 hours (stored) to be measured thereafter identical to the spiked samples.

5. Analysis / Results

5.1. Time of realization

The study was performed between 11th of September and 4th of October 2018.

5.2. Data

A total of 100 patient sample data, which fulfilled the inclusion criteria, were included in the statistical analysis.

Another 5 data sets had to be excluded (shown in 4.4.).

5.3. Analysis of system accuracy in accordance with DIN EN ISO 15197:2015

The analysis of data for the proof of system accuracy was done in accordance with the instructions given in the DIN EN ISO 15197:2015 [2]. The test measurements were compared with the reference values determined using the YSI 2300 STAT PLUS system.

For samples with concentrations < 100 mg/dl the deviation of test values from the mean of the reference values were calculated in mg/dl. For samples with glucose concentrations ≥ 100 mg/dl the percentage deviation from the mean of the reference values were calculated.

The following parameters were calculated:

Drift in %	deviation of the reference values from YSI values in %
YSI average	mean of reference values in mg/dl
Dev. GC to YSI in mg/dl	average deviation of the measured values to the reference values in mg/dl
Dev. GC to YSI in %	average deviation of the measured values to the reference values in %

Result of system accuracy of On Call Sure for glucose concentrations < 100 mg/dL

	within ± 5mg/dL	within ± 10 mg/dL	within ± 15 mg/dL
Lot 1	29 / 62 (46.8 %)	59 / 62 (95.2 %)	62 / 62 (100 %)
Lot 2	31 / 62 (50.0 %)	60 / 62 (96.8 %)	62 / 62 (100 %)
Lot 3	31 / 62 (50.0 %)	61 / 62 (98.4 %)	62 / 62 (100 %)
Lot 1, 2 and 3 in summary	91 / 186 (48.9 %)	180 / 186 (96.8 %)	186 / 186 (100 %)

Result of system accuracy of On Call Sure for glucose concentrations ≥ 100 mg/dL

	within ± 5 %	within ± 10 %	within ± 15 %
Lot 1	87 / 138 (63.0 %)	131 / 138 (94.9 %)	137 / 138 (99.3 %)
Lot 2	99 / 138 (71.7 %)	131 / 138 (94.9 %)	138 / 138 (100 %)
Lot 3	86 / 138 (62.3 %)	131 / 138 (94.9 %)	137 / 138 (99.3 %)
Lots 1, 2 and 3	272 / 414 (65.7 %)	393 / 414 (94.9 %)	412 / 414 (99.5 %)

System accuracy of On Call Sure for combined glucose concentrations

	within ± 15 mg/dL or 15 %
Lot 1	199 / 200 (99.5 %)
Lot 2	200 / 200 (100 %)
Lot 3	199 / 200 (99.5 %)
Lots 1, 2 and 3	598 / 600 (99.7 %)

In all three test strip lots more than 95 % of the samples had a deviation to the reference method of less than 15 mg/dl in samples with blood glucose levels of < 100 mg/dl and thus fulfilled the criteria of the quality norm DIN EN 15197:2015. In samples with blood glucose concentrations of ≥ 100 mg/dl less than 5 % of the readings in all three test strip lots revealed deviations from the reference method of less than 15 %. The On Call Sure in all three test strip lots fulfilled the criteria of the quality norm.

A regression analysis was performed according to Passing and Bablok [4] and the readings are shown in the Error Grid Diagram.

Figure 1: Error Grid Analysis of test Lot 1 (1790007), Mean values of On Call Sure vs. mean values of the YSI 2300 Stat Plus




Figure 2: Graphic presentation of the differences between the On Call Sure (Lot 1) and reference measurements, as a function of the glucose concentration




Figure 3: Error Grid Analysis of test Lot 2 (1790008), Mean values of On Call Sure vs. mean values of the YSI 2300 Stat Plus




Figure 4: Graphic presentation of the differences between the One Call Sure (Lot 2) and reference measurements, as a function of the glucose concentration




Figure 5: Error Grid Analysis of test Lot 3 (1790009), Mean values of On Call Sure vs. mean values of the YSI 2300 Stat Plus





Figure 6: Graphic presentation of the differences between the On Call Sure (Lot 3) and reference measurements, as a function of the glucose concentration



The readings of all three test lots were within the region A and B of Parkes Error Grid.

6. Summary and conclusions

The criteria for the system accuracy according to the standard DIN EN ISO 15197:2015 were fulfilled for the blood glucose monitoring system On Call Sure:

- 100 % of the performed measurements were within the zone A and B of the Consensus Error Grid [5].
- According to the results of system accuracy 99.7 % of all measurements were within the required range.
- For the On Call Sure system, the criteria of the DIN EN ISO 15197:2015 were met by the study data.

References

1. Testsysteme für die In-vitro-Diagnostik – Anforderungen an Blutzuckermesssysteme zur Eigenanwendung beim Diabetes mellitus (DIN EN ISO 15197:2003)
2. Testsysteme für die In-vitro-Diagnostik - Anforderungen an Blutzuckermesssysteme zur Eigenanwendung bei Diabetes mellitus (DIN EN ISO 15197:2015).
3. Bland J.M. and Altman D.G.: Statistical methods for assessing agreement between two methods of clinical measurement. Lancet 1986, 307-310.
4. Passing H. and Bablok W.: A new biometrical procedure for testing the equality of measurements from two different analytical methods. J Clin Chem Clin Biochem Vol. 21, 1983: 709-720.
5. Parkes J.L., Slatin S.L., Pardo S., Ginsberg B.H.: A New Consensus Error Grid to Evaluate the Clinical Significance of Inaccuracies in the Measurement of Blood Glucose. Diabetes Care 2000. 23:1143–1148.

Signatures

The undersigned has reviewed the format and content of this report.



Dr. Eckhard Salzsieder
Studienleiter und Institutsdirektor
Institut für Diabetes „Gerhardt Katsch“ Karlsburg e.V.
Greifswalder Str. 11 E
17495 Karlsburg